

.

From: Nassif, Julianne (DPH)
Sent: Friday, December 02, 2011 12:00 PM
To: Servizio, Paul (DPH)
Subject: RE: morphine samples

Do a little research i.e, models, costs etc maybe we can do it.

From: Servizio, Paul (DPH)
Sent: Friday, December 02, 2011 11:45 AM
To: Nassif, Julianne (DPH)
Subject: RE: morphine samples

It would be good to buy another turbovap 96 The one we presently have works but is taking 2 to 3 hours to blow down instead of 45 minutes,. Pete John and I were checking out the pressure and fittings today. It may work better now, but a backup is recommended.
Paul

From: Nassif, Julianne (DPH)
Sent: Friday, December 02, 2011 10:33 AM
To: Clemmer, Jill (DPH); Piro, Peter (DPH); O'Brien, Elisabeth (DPH); Servizio, Paul (DPH)
Subject: morphine samples
Importance: High

After discussing in more detail with Peter, I think it is best that we change the certificates to reflect the higher reporting limit. Once the new certificates are complete, I will contact Sgt Rouvalis and ADA Walsh to explain.
Thanks, Julie

Julianne Nassif
Director of Analytical Chemistry
William A. Hinton State Laboratory
Massachusetts Department of Public Health
305 South Street Boston, MA 02130
617.983.6651 (voice) 617.983.6662 (fax)